

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092221\
 Data File : VU044759.D
 Acq On : 22 Sep 2021 15:27
 Operator : SY/MD
 Sample : M3814-05MS
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GB6W6MS

Quant Time: Sep 23 03:04:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090921WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 23 02:59:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	123443	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	125860	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	55535	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	32478	4.459	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.200%
7) Chloroethane-d5	1.916	69	31636	4.256	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.200%
11) 1,1-Dichloroethene-d2	2.571	65	13791	3.749	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.000%
20) 2-Butanone-d5	4.639	46	136154	44.593	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.180%
24) Chloroform-d	5.070	84	70651	4.612	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.200%
26) 1,2-Dichloroethane-d4	5.710	65	42941	4.701	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.000%
32) Benzene-d6	5.735	84	133202	4.151	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.000%
36) 1,2-Dichloropropane-d6	6.697	67	47820	4.537	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.800%
41) Toluene-d8	7.902	98	105205	3.766	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.400%
43) trans-1,3-Dichloroprop...	8.182	79	16116	4.338	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.800%
46) 2-Hexanone-d5	8.642	63	103604	43.390	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.780%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	42556	4.554	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.000%
66) 1,2-Dichlorobenzene-d4	12.198	152	42440	4.631	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.600%
Target Compounds						Qvalue
13) Acetone	2.652	43	4090	2.057	ug/L	87
42) Toluene	7.976	91	5062	0.130	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092221\
Data File : VU044759.D
Acq On : 22 Sep 2021 15:27
Operator : SY/MD
Sample : M3814-05MS
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GB6W6MS

Quant Time: Sep 23 03:04:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090921WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 23 02:59:45 2021
Response via : Initial Calibration

