


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092221\
Data File : VU044769.D
Acq On : 22 Sep 2021 19:28
Operator : SY/MD
Sample : M3814-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampled :
GB6X5

Quant Time: Sep 23 03:06:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090921WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 23 02:59:45 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

MMDadoda
9/24/2021 1:35:35 PM



Quantitation Report (Qedit)




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092221\
 Data File : VU044769.D
 Acq On : 22 Sep 2021 19:28
 Operator : SY/MD
 Sample : M3814-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GB6X5

Quant Time: Sep 23 03:06:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090921WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 23 02:59:45 2021
 Response via : Initial Calibration

Manual Integrations
 APPROVED

MMDadoda
 9/24/2021 1:35:35 PM



TIC: VU044769.D\data.ms

(13) Acetone (T)

2.652min (+ 0.003) 4.15 ug/L

response 8000

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	21.20	10.10
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092221\
 Data File : VU044769.D
 Acq On : 22 Sep 2021 19:28
 Operator : SY/MD
 Sample : M3814-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GB6X5

Quant Time: Sep 23 03:06:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090921WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 23 02:59:45 2021
 Response via : Initial Calibration

Manual Integrations
 APPROVED

MMDadoda
 9/24/2021 1:35:35 PM



TIC: VU044769.D\data.ms

(13) Acetone (T)

2.652min (+ 0.003) 4.54 ug/L m

response 8742

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	21.20	9.24
0.00	0.00	0.00
0.00	0.00	0.00

MD
 9/28/21

Quantitation Report (QT Reviewed)

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092221\
 Data File : VU044769.D
 Acq On : 22 Sep 2021 19:28
 Operator : SY/MD
 Sample : M3814-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 Client Sample ID :
 GB6X5

Manual Integrations
 APPROVED

MMDadoda
 9/24/2021 1:35:35 PM

Quant Time: Sep 23 03:06:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090921WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 23 02:59:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	119566	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	122194	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	54440	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	31774	4.504	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery =	90.000%		
7) Chloroethane-d5	1.916	69	30417	4.225	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery =	84.400%		
11) 1,1-Dichloroethene-d2	2.571	65	13961	3.918	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery =	78.400%		
20) 2-Butanone-d5	4.639	46	143310	48.459	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery =	96.920%		
24) Chloroform-d	5.070	84	72216	4.867	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery =	97.400%		
26) 1,2-Dichloroethane-d4	5.710	65	43785	4.948	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery =	99.000%		
32) Benzene-d6	5.735	84	130113	4.176	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery =	83.600%		
36) 1,2-Dichloropropane-d6	6.697	67	47245	4.617	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery =	92.400%		
41) Toluene-d8	7.902	98	103856	3.829	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery =	76.600%		
43) trans-1,3-Dichloroprop...	8.185	79	15442	4.281	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery =	85.600%		
46) 2-Hexanone-d5	8.645	63	107047	46.177	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery =	92.360%		
56) 1,1,2,2-Tetrachloroeth...	10.761	84	43205	4.762	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery =	95.200%		
66) 1,2-Dichlorobenzene-d4	12.198	152	41732	4.646	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery =	93.000%		
Target Compounds						
13) Acetone	2.652	43	8742m	4.539	ug/L	Qvalue
42) Toluene	7.973	91	4270	0.113	ug/L	97

7/170
 9/28/21

(#) = qualifier out of range (m) = manual integration (+) = signals summed