

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092222\
 Data File : VU050901.D
 Acq On : 22 Sep 2022 16:27
 Operator : JC/MD
 Sample : N4747-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E45A0

Quant Time: Sep 23 01:11:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 23 01:08:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	381513	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	377094	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	138950	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	136986	34.656	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	69.320%	
7) Chloroethane-d5	1.906	69	120125	42.672	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	85.340%	
11) 1,1-Dichloroethene-d2	2.565	63	180341	29.321	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	58.640%#	
21) 2-Butanone-d5	4.617	46	300467	117.744	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	117.740%	
24) Chloroform-d	5.060	84	277965	44.873	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	89.740%	
26) 1,2-Dichloroethane-d4	5.700	65	183735	48.130	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.260%	
32) Benzene-d6	5.726	84	525732	42.327	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	84.660%	
36) 1,2-Dichloropropane-d6	6.690	67	190137	45.351	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	90.700%	
41) Toluene-d8	7.896	98	413191	38.057	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	76.120%#	
43) trans-1,3-Dichloroprop...	8.179	79	73683	40.399	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	80.800%	
47) 2-Hexanone-d5	8.632	63	169314	97.997	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	98.000%	
56) 1,1,2,2-Tetrachloroeth...	10.755	84	298219	46.655	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	93.320%	
66) 1,2-Dichlorobenzene-d4	12.192	152	145335	49.956	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	99.920%	
Target Compounds						
					Qvalue	
34) Trichloroethene	6.542	95	11900	3.692	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092222\
Data File : VU050901.D
Acq On : 22 Sep 2022 16:27
Operator : JC/MD
Sample : N4747-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E45A0

Quant Time: Sep 23 01:11:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 23 01:08:30 2022
Response via : Initial Calibration

