

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092223\
 Data File : VU055463.D
 Acq On : 22 Sep 2023 16:38
 Operator : MD/SY
 Sample : 04562-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DCJK7

Quant Time: Sep 25 01:11:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Sep 25 01:08:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	315633	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	331994	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	156976	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	107572	3.372	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	67.400%
7) Chloroethane-d5	1.911	69	104926	3.732	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	74.600%
11) 1,1-Dichloroethene-d2	2.563	65	48285	3.677	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.600%
20) 2-Butanone-d5	4.631	46	307257	53.325	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.640%
24) Chloroform-d	5.058	84	265229	4.597	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.000%
26) 1,2-Dichloroethane-d4	5.698	65	137917	4.969	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.400%
32) Benzene-d6	5.724	84	500758	4.312	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.200%
36) 1,2-Dichloropropane-d6	6.689	67	160403	4.623	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.400%
41) Toluene-d8	7.894	98	422578	3.993	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.800%
43) trans-1,3-Dichloroprop...	8.177	79	54066	4.440	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.800%
46) 2-Hexanone-d5	8.631	63	156197	42.175	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.340%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	123407	4.746	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	155054	4.894	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.800%
Target Compounds						Qvalue
13) Acetone	2.650	43	25174	7.664	ug/L	96
34) Trichloroethene	6.541	95	75785	2.804	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092223\
Data File : VU055463.D
Acq On : 22 Sep 2023 16:38
Operator : MD/SY
Sample : 04562-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DCJK7

Quant Time: Sep 25 01:11:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Sep 25 01:08:17 2023
Response via : Initial Calibration

