

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092318\
 Data File : VU027049.D
 Acq On : 23 Sep 2018 04:57
 Operator : MD/SY
 Sample : J4958-16
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 91 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-TT-AOC22-MW10-20180915

Quant Time: Sep 24 06:36:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	94354	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	156572	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	149653	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	70990	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	97071	51.12	ug/l	0.00
Spiked Amount			Recovery	=	102.24%	
35) Dibromofluoromethane	4.89	113	67180	49.50	ug/l	0.00
Spiked Amount			Recovery	=	99.00%	
50) Toluene-d8	7.57	98	252633	48.84	ug/l	0.00
Spiked Amount			Recovery	=	97.68%	
62) 4-Bromofluorobenzene	10.31	95	80766	45.51	ug/l	0.00
Spiked Amount			Recovery	=	91.02%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092318\
Data File : VU027049.D
Acq On : 23 Sep 2018 04:57
Operator : MD/SY
Sample : J4958-16
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 91 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-TT-AOC22-MW10-20180915

Quant Time: Sep 24 06:36:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

