

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092318\
 Data File : VU027050.D
 Acq On : 23 Sep 2018 05:20
 Operator : MD/SY
 Sample : J4963-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 92 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SA-PZ123I-20180914

Quant Time: Sep 24 06:38:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	94043	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	157843	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	152160	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	70372	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	96407	50.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.86%	
35) Dibromofluoromethane	4.88	113	66817	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
50) Toluene-d8	7.57	98	254842	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
62) 4-Bromofluorobenzene	10.31	95	81624	45.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.24%	
Target Compounds						
6) Chloroethane	1.70	64	4385	3.473	ug/l	91
12) 1,1-Dichloroethene	2.29	96	4441	3.419	ug/l	93
24) 1,1-Dichloroethane	3.45	63	85105	27.423	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092318\
Data File : VU027050.D
Acq On : 23 Sep 2018 05:20
Operator : MD/SY
Sample : J4963-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 92 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SA-PZ123I-20180914

Quant Time: Sep 24 06:38:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

