

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092318\
 Data File : VU027055.D
 Acq On : 23 Sep 2018 07:16
 Operator : MD/SY
 Sample : J4963-09
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 97 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SA-PZ127I-20180914

Quant Time: Sep 24 06:49:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	93248	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	160155	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	154616	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	71178	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	98585	52.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.06%	
35) Dibromofluoromethane	4.88	113	69126	49.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.58%	
50) Toluene-d8	7.57	98	249443	47.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.30%	
62) 4-Bromofluorobenzene	10.31	95	83388	45.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.86%	
Target Compounds						
64) Tetrachloroethene	8.23	164	4048	3.398	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092318\
Data File : VU027055.D
Acq On : 23 Sep 2018 07:16
Operator : MD/SY
Sample : J4963-09
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 97 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SA-PZ127I-20180914

Quant Time: Sep 24 06:49:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

