

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092318\
 Data File : VU027071.D
 Acq On : 23 Sep 2018 14:56
 Operator : MD/SY
 Sample : J4954-05
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 113 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SA-PZ124-20180914

Quant Time: Sep 24 08:34:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	104840	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	178135	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	170680	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	77224	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	110430	52.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.66%	
35) Dibromofluoromethane	4.89	113	75492	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
50) Toluene-d8	7.57	98	285831	48.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.14%	
62) 4-Bromofluorobenzene	10.31	95	92804	45.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.92%	
Target Compounds						
6) Chloroethane	1.70	64	4736	3.365	ug/l	96
12) 1,1-Dichloroethene	2.29	96	2960	2.044	ug/l	95
24) 1,1-Dichloroethane	3.45	63	59913	17.317	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092318\
Data File : VU027071.D
Acq On : 23 Sep 2018 14:56
Operator : MD/SY
Sample : J4954-05
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 113 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SA-PZ124-20180914

Quant Time: Sep 24 08:34:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

