

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092318\
 Data File : VU027074.D
 Acq On : 23 Sep 2018 16:05
 Operator : MD/SY
 Sample : J4954-14
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 116 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SA-PZ118I-20180914

Quant Time: Sep 24 08:40:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	101579	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	174727	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	165668	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	74201	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	106079	51.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.78%	
35) Dibromofluoromethane	4.88	113	73601	48.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.20%	
50) Toluene-d8	7.57	98	277793	48.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.26%	
62) 4-Bromofluorobenzene	10.31	95	89326	45.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.20%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092318\
Data File : VU027074.D
Acq On : 23 Sep 2018 16:05
Operator : MD/SY
Sample : J4954-14
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 116 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SA-PZ118I-20180914

Quant Time: Sep 24 08:40:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

