

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092319\
 Data File : VU034755.D
 Acq On : 24 Sep 2019 00:03
 Operator : JC/SP
 Sample : K4932-12
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFDK0

Quant Time: Sep 24 05:42:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091919WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Sep 24 03:36:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	525710	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	492786	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.46	152	239585	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	209338	38.90	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.80%
7) Chloroethane-d5	1.67	69	175016	42.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.16%
11) 1,1-Dichloroethene-d2	2.26	63	297771	32.67	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.34%
21) 2-Butanone-d5	4.15	46	457085	110.81	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	110.81%
24) Chloroform-d	4.62	84	370320	47.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.66%
26) 1,2-Dichloroethane-d4	5.29	65	274207	47.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.34%
32) Benzene-d6	5.32	84	754835	49.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.84%
36) 1,2-Dichloropropane-d6	6.31	67	267787	50.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.32%
41) Toluene-d8	7.54	98	716929	49.34	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.68%
43) trans-1,3-Dichloropropene-	7.83	79	116821	45.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.38%
47) 2-Hexanone-d5	8.29	63	304019	113.38	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	113.38%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	357473	48.26	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.52%
64) 1,2-Dichlorobenzene-d4	11.84	152	257426	51.24	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.48%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Acetone	2.31	43	203526	42.418	ug/L	100
15) Methyl Acetate	2.61	43	6032	0.941	ug/L	# 77
16) Methylene chloride	2.68	84	64831	14.880	ug/L	93
22) 2-Butanone	4.24	43	57954	10.153	ug/L	100
40) 4-Methyl-2-pentanone	7.44	43	29755	3.063	ug/L	99
42) Toluene	7.62	91	44236	2.395	ug/L	91
51) Chlorobenzene	9.10	112	16763	1.493	ug/L	97
52) Ethylbenzene	9.23	91	93781	4.509	ug/L	98
53) m,p-Xylene	9.35	106	64108	8.702	ug/L	97
54) o-xylene	9.76	106	53590	7.355	ug/L	100
56) Isopropylbenzene	10.14	105	54938	2.796	ug/L	97
63) 1,4-Dichlorobenzene	11.48	146	7440	0.880	ug/L	90
65) 1,2-Dichlorobenzene	11.85	146	26237	3.130	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092319\
Data File : VU034755.D
Acq On : 24 Sep 2019 00:03
Operator : JC/SP
Sample : K4932-12
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFDK0

Quant Time: Sep 24 05:42:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091919WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 24 03:36:55 2019
Response via : Initial Calibration

