

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092320\
 Data File : VU040255.D
 Acq On : 23 Sep 2020 11:25
 Operator : SY/MD
 Sample : VU0923WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK12

Quant Time: Sep 29 02:08:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Sep 24 02:21:14 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	158827	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	167058	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	70498	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	44882	5.18	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.60%
7) Chloroethane-d5	1.92	69	41118	5.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.80%
11) 1,1-Dichloroethene-d2	2.58	63	74841	3.36	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.20%
20) 2-Butanone-d5	4.65	46	170003	49.01	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.02%
24) Chloroform-d	5.08	84	96142	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.00%
26) 1,2-Dichloroethane-d4	5.72	65	59506	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.20%
32) Benzene-d6	5.74	84	186117	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.40%
36) 1,2-Dichloropropane-d6	6.70	67	62246	4.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.80%
41) Toluene-d8	7.91	98	147865	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.00%
43) trans-1,3-Dichloropropene-	8.19	79	23056	3.90	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.00%
46) 2-Hexanone-d5	8.64	63	115897	41.43	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.86%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	49675	4.47	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	57949	5.08	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.60%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	2005	0.125 ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092320\
Data File : VU040255.D
Acq On : 23 Sep 2020 11:25
Operator : SY/MD
Sample : VU0923WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK12

Quant Time: Sep 29 02:08:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Sep 24 02:21:14 2020
Response via : Initial Calibration

