

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092321\
 Data File : VU044787.D
 Acq On : 23 Sep 2021 16:13
 Operator : SY/MD
 Sample : M3873-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EW7X8

Quant Time: Sep 24 05:38:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 24 05:26:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	265437	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	260261	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	112947	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	133189	45.836	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	91.680%		
7) Chloroethane-d5	1.912	69	107097	46.511	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	93.020%		
11) 1,1-Dichloroethene-d2	2.572	63	168174	34.464	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	68.920%		
21) 2-Butanone-d5	4.633	46	238339	103.751	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	103.750%		
24) Chloroform-d	5.070	84	210000	45.758	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	91.520%		
26) 1,2-Dichloroethane-d4	5.706	65	146533	47.137	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.280%		
32) Benzene-d6	5.732	84	427993	47.392	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.780%		
36) 1,2-Dichloropropane-d6	6.697	67	141774	48.145	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	96.300%		
41) Toluene-d8	7.903	98	353300	45.332	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	90.660%		
43) trans-1,3-Dichloroprop...	8.182	79	59274	45.251	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	90.500%		
47) 2-Hexanone-d5	8.639	63	166399	101.969	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	101.970%		
56) 1,1,2,2-Tetrachloroeth...	10.761	84	207354	45.220	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	90.440%		
66) 1,2-Dichlorobenzene-d4	12.198	152	125165	50.921	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	101.840%		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.575	96	2018	0.982	ug/L	# 1
17) trans-1,2-Dichloroethene	3.359	96	12987	5.998	ug/L	98
20) cis-1,2-Dichloroethene	4.671	96	610658	255.932	ug/L	97
34) Trichloroethene	6.546	95	276361	119.196	ug/L	99
42) Toluene	7.973	91	5648	0.576	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092321\
Data File : VU044787.D
Acq On : 23 Sep 2021 16:13
Operator : SY/MD
Sample : M3873-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EW7X8

Quant Time: Sep 24 05:38:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092321WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 24 05:26:38 2021
Response via : Initial Calibration

