

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092321\
 Data File : VU044786.D
 Acq On : 23 Sep 2021 15:50
 Operator : SY/MD
 Sample : M3873-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EW7X7

Quant Time: Sep 24 05:38:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 24 05:26:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	272856	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	261662	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	117968	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	132324	44.300	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	88.600%		
7) Chloroethane-d5	1.912	69	103414	43.690	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	87.380%		
11) 1,1-Dichloroethene-d2	2.572	63	164370	32.768	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	65.540%		
21) 2-Butanone-d5	4.629	46	235020	99.525	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	99.520%		
24) Chloroform-d	5.070	84	202923	43.014	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	86.020%		
26) 1,2-Dichloroethane-d4	5.707	65	144411	45.192	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	90.380%		
32) Benzene-d6	5.732	84	416273	45.847	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	91.700%		
36) 1,2-Dichloropropane-d6	6.697	67	137298	46.376	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	92.760%		
41) Toluene-d8	7.903	98	344539	43.971	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	87.940%		
43) trans-1,3-Dichloroprop...	8.182	79	57507	43.667	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	87.340%		
47) 2-Hexanone-d5	8.636	63	159493	97.213	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	97.210%		
56) 1,1,2,2-Tetrachloroeth...	10.761	84	201947	43.805	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	87.620%		
66) 1,2-Dichlorobenzene-d4	12.198	152	123302	48.028	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	96.060%		
Target Compounds						Qvalue
3) Chloromethane	1.520	50	2147	0.659	ug/L	96
5) Vinyl chloride	1.607	62	390055	122.256	ug/L	99
12) 1,1-Dichloroethene	2.578	96	2637	1.248	ug/L #	1
17) trans-1,2-Dichloroethene	3.356	96	22985	10.326	ug/L	95
20) cis-1,2-Dichloroethene	4.671	96	899244	366.634	ug/L	98
29) Cyclohexane	5.391	56	31518	8.013	ug/L	98
33) Benzene	5.780	78	29061	2.994	ug/L	100
34) Trichloroethene	6.546	95	10490	4.500	ug/L	97
42) Toluene	7.973	91	5746	0.583	ug/L	83


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092321\
Data File : VU044786.D
Acq On : 23 Sep 2021 15:50
Operator : SY/MD
Sample : M3873-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EW7X7

Quant Time: Sep 24 05:38:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092321WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 24 05:26:38 2021
Response via : Initial Calibration









#3
Chloromethane
Concen: 0.659 ug/L
RT: 1.520 min Scan# 56
Delta R.T. -0.000 min
Lab File: VU044786.D
Acq: 23 Sep 2021 15:50




Instrument :
MSVOA_U
ClientSampled :
EW7X7


Tgt Ion: 50 Resp: 2147
Ion Ratio Lower Upper
50 100
52 35.4 23.0 42.8



#5
Vinyl chloride
Concen: 122.256 ug/L
RT: 1.607 min Scan# 83
Delta R.T. -0.000 min
Lab File: VU044786.D
Acq: 23 Sep 2021 15:50

Tgt Ion: 62 Resp: 390055
Ion Ratio Lower Upper
62 100
64 32.7 23.4 43.6









#12
1,1-Dichloroethene
Concen: 1.248 ug/L
RT: 2.578 min Scan# 385
Delta R.T. -0.003 min
Lab File: VU044786.D
Acq: 23 Sep 2021 15:50




Instrument :
MSVOA_U
ClientSampleId :
EW7X7


Tgt Ion: 96 Resp: 2637
Ion Ratio Lower Upper
96 100
61 809.8 131.3 243.9#
63 5759.0 102.5 190.3#



#17
trans-1,2-Dichloroethene
Concen: 10.326 ug/L
RT: 3.356 min Scan# 627
Delta R.T. -0.000 min
Lab File: VU044786.D
Acq: 23 Sep 2021 15:50

Tgt Ion: 96 Resp: 22985
Ion Ratio Lower Upper
96 100
61 157.9 105.1 195.3
98 62.5 43.0 79.8









#20
cis-1,2-Dichloroethene
Concen: 366.634 ug/L
RT: 4.671 min Scan# 1036
Delta R.T. -0.003 min
Lab File: VU044786.D
Acq: 23 Sep 2021 15:50




Instrument :
MSVOA_U
ClientSampled :
EW7X7


Tgt Ion:	96	Resp:	899244
Ion	Ratio	Lower	Upper
96	100		
61	140.8	100.2	186.0
68	0.0	0.0	0.0



#29
Cyclohexane
Concen: 8.013 ug/L
RT: 5.391 min Scan# 1260
Delta R.T. -0.003 min
Lab File: VU044786.D
Acq: 23 Sep 2021 15:50





Tgt Ion:	56	Resp:	31518
Ion	Ratio	Lower	Upper
56	100		
69	28.7	23.8	35.8
84	85.2	66.8	100.2








#33
Benzene
Concen: 2.994 ug/L
RT: 5.780 min Scan# 1381
Delta R.T. -0.000 min
Lab File: VU044786.D
Acq: 23 Sep 2021 15:50
Tgt Ion: 78 Resp: 29061


Instrument :
MSVOA_U
ClientSampled :
EW7X7



#34
Trichloroethene
Concen: 4.500 ug/L
RT: 6.546 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VU044786.D
Acq: 23 Sep 2021 15:50
Tgt Ion: 95 Resp: 10490

Ion	Ratio	Lower	Upper
95	100		
97	65.7	44.4	82.4
132	88.3	62.9	116.9
130	87.9	65.0	120.8





#42
Toluene
Concen: 0.583 ug/L
RT: 7.973 min Scan# 2063
Delta R.T. -0.000 min
Lab File: VU044786.D
Acq: 23 Sep 2021 15:50

Instrument :
MSVOA_U
ClientSampleId :
EW7X7

Tgt Ion: 91 Resp: 5746
Ion Ratio Lower Upper
91 100
92 70.1 40.5 75.1

