

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092418\
 Data File : VU027102.D
 Acq On : 24 Sep 2018 12:39
 Operator : MD/SY
 Sample : J4969-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SA-PZ144I-20180916

Quant Time: Sep 25 02:08:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	94716	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	161570	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	156116	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	71635	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	100637	52.79	ug/l	0.00
Spiked Amount			Recovery	=	105.58%	
35) Dibromofluoromethane	4.88	113	68461	48.88	ug/l	0.00
Spiked Amount			Recovery	=	97.76%	
50) Toluene-d8	7.57	98	260513	48.81	ug/l	0.00
Spiked Amount			Recovery	=	97.62%	
62) 4-Bromofluorobenzene	10.31	95	83640	45.67	ug/l	0.00
Spiked Amount			Recovery	=	91.34%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092418\
Data File : VU027102.D
Acq On : 24 Sep 2018 12:39
Operator : MD/SY
Sample : J4969-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SA-PZ144I-20180916

Quant Time: Sep 25 02:08:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

