

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092418\
 Data File : VU027112.D
 Acq On : 24 Sep 2018 16:31
 Operator : MD/SY
 Sample : J4969-15
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SA-MW129S-20180917

Quant Time: Sep 25 02:47:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	92239	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	156701	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	147868	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	64372	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	96664	52.07	ug/l	0.00
Spiked Amount			Recovery	=	104.14%	
35) Dibromofluoromethane	4.89	113	66437	48.91	ug/l	0.00
Spiked Amount			Recovery	=	97.82%	
50) Toluene-d8	7.57	98	251342	48.55	ug/l	0.00
Spiked Amount			Recovery	=	97.10%	
62) 4-Bromofluorobenzene	10.31	95	78307	44.09	ug/l	0.00
Spiked Amount			Recovery	=	88.18%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092418\
Data File : VU027112.D
Acq On : 24 Sep 2018 16:31
Operator : MD/SY
Sample : J4969-15
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SA-MW129S-20180917

Quant Time: Sep 25 02:47:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

