

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU092419\
Quantitation Report (QT Reviewed)


Data File : VU034769.D
Acq On : 24 Sep 2019 14:27
Operator : JC/SP
Sample : K4984-08 10X
Misc : 6.22g/5.0ml/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEDL2

Manual Integrations
APPROVED

MMDadoda
9/25/2019 10:06:42 AM

Quant Time: Sep 25 03:47:24 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091919WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 25 02:59:31 2019
Response via : Initial Calibration



Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU092419\
Quantitation Report (Qedit)




Data File : VU034769.D
Acq On : 24 Sep 2019 14:27
Operator : JC/SP
Sample : K4984-08 10X
Misc : 6.22g/5.0ml/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEDL2

Manual Integrations
APPROVED

MMDadoda
9/25/2019 10:06:42 AM

Quant Time: Sep 25 03:02:36 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091919WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 25 02:59:31 2019
Response via : Initial Calibration



TIC: VU034769.D

(7) Chloroethane-d5 (S)
1.910min (+0.241) 0.11ug/L
response 379

Ion	Exp%	Act%
69.00	100	100
71.00	31.80	27.44
51.00	32.70	30.34
0.00	0.00	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU092419\
Quantitation Report (Qedit)




Data File : VU034769.D
Acq On : 24 Sep 2019 14:27
Operator : JC/SP
Sample : K4984-08 10X
Misc : 6.22g/5.0ml/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEDL2

Manual Integrations
APPROVED

MMDadoda
9/25/2019 10:06:42 AM

Quant Time: Sep 25 03:02:36 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091919WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 25 02:59:31 2019
Response via : Initial Calibration



TIC: VU034769.D

(7) Chloroethane-d5 (S)

1.656min (-0.013) 45.61ug/L m

response 161828

Ion	Exp%	Act%
69.00	100	100
71.00	31.80	0.06#
51.00	32.70	0.07#
0.00	0.00	0.00

7 mQ
913 = 119

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU092419\
Quantitation Report (QT Reviewed)

Data File : VU034769.D
Acq On : 24 Sep 2019 14:27
Operator : JC/SP
Sample : K4984-08 10X
Misc : 6.22g/5.0ml/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEDL2

Manual Integrations
APPROVED

MMDadoda
9/25/2019 10:06:42 AM

Quant Time: Sep 25 03:47:24 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091919WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 25 02:59:31 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	448476	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	440531	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.46	152	201433	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	185436	40.39	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.78%
7) Chloroethane-d5	1.66	69	161828m	45.61	ug/L	-0.01
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.22%
11) 1,1-Dichloroethene-d2	2.25	63	253060	32.55	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.10%
21) 2-Butanone-d5	4.15	46	381200	108.33	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.33%
24) Chloroform-d	4.62	84	325693	48.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.60%
26) 1,2-Dichloroethane-d4	5.28	65	235521	48.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.00%
32) Benzene-d6	5.32	84	691335	50.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.26%
36) 1,2-Dichloropropane-d6	6.31	67	246944	51.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.48%
41) Toluene-d8	7.54	98	672493	51.77	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.54%
43) trans-1,3-Dichloropropene-	7.83	79	111119	48.61	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.22%
47) 2-Hexanone-d5	8.29	63	261387	109.05	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.05%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	350138	52.87	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.74%
64) 1,2-Dichlorobenzene-d4	11.84	152	238729	56.52	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	113.04%

7 ml
09/30/19

Target Compounds					Qvalue
35) Methylcyclohexane	6.39	83	9376	1.464 ug/L	91
46) Tetrachloroethene	8.21	164	7019	2.644 ug/L	98
52) Ethylbenzene	9.23	91	20184	1.086 ug/L	98
53) m,p-Xylene	9.36	106	17542	2.664 ug/L	88
56) Isopropylbenzene	10.14	105	21906	1.247 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed