

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092420\
 Data File : VU040267.D
 Acq On : 24 Sep 2020 12:42
 Operator : SY/MD
 Sample : VU0924WBL02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK202

Quant Time: Sep 25 09:53:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMULM092420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 25 09:34:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	400870	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	393974	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.81	152	180622	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	151450	45.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.86%
7) Chloroethane-d5	1.92	69	117299	46.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.86%
11) 1,1-Dichloroethene-d2	2.58	63	228705	35.30	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.60%
21) 2-Butanone-d5	4.65	46	292652	103.62	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.62%
24) Chloroform-d	5.08	84	273755	46.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.38%
26) 1,2-Dichloroethane-d4	5.72	65	198876	47.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.96%
32) Benzene-d6	5.75	84	535887	48.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.08%
36) 1,2-Dichloropropane-d6	6.70	67	181072	47.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.88%
41) Toluene-d8	7.91	98	469058	46.62	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.24%
43) trans-1,3-Dichloropropene-	8.19	79	89886	46.80	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.60%
47) 2-Hexanone-d5	8.64	63	176950	105.74	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.74%
56) 1,1,2,2-Tetrachloroethane-	10.75	84	250005	46.77	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.54%
66) 1,2-Dichlorobenzene-d4	12.19	152	178659	51.41	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.82%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092420\
Data File : VU040267.D
Acq On : 24 Sep 2020 12:42
Operator : SY/MD
Sample : VU0924WBL02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK202

Quant Time: Sep 25 09:53:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SFAMULM092420WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 25 09:34:00 2020
Response via : Initial Calibration

