

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092425\
 Data File : VU063637.D
 Acq On : 24 Sep 2025 13:19
 Operator : MD/SY
 Sample : VU0924WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK025

Quant Time: Sep 25 04:18:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM092425.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Sep 25 04:13:39 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	7079	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.418	117	9393	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.825	152	3204	0.500	ug/L	0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.590	65	2124	0.447	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	90.000%
4) 1,2-Dichloroethane-d4	5.709	65	1188	0.421	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	84.000%
7) 1,2-Dichloropropane-d6	6.687	67	1238	0.503	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	100.000%
8) Toluene-d8	7.887	98	3579	0.454	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	90.000%
10) 1,1,2,2-Tetrachloroeth...	10.745	84	1032	0.471	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	94.000%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092425\
Data File : VU063637.D
Acq On : 24 Sep 2025 13:19
Operator : MD/SY
Sample : VU0924WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK025

Quant Time: Sep 25 04:18:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM092425.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Sep 25 04:13:39 2025
Response via : Initial Calibration

