

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
 Data File : VU044826.D
 Acq On : 24 Sep 2021 22:03
 Operator : SY/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 IBLK

Quant Time: Sep 25 01:21:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 01:18:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	486	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.259	114	677	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	939	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	1405	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.713	65	712	83.942	ug/l	0.00
Spiked Amount	50.000		Recovery	= 167.880%		
35) Dibromofluoromethane	5.295	113	478	102.921	ug/l	0.00
Spiked Amount	50.000		Recovery	= 205.840%		
50) Toluene-d8	7.902	98	1333	75.995	ug/l	0.00
Spiked Amount	50.000		Recovery	= 152.000%		
62) 4-Bromofluorobenzene	10.639	95	1363	209.211	ug/l	0.00
Spiked Amount	50.000		Recovery	= 418.420%		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
Data File : VU044826.D
Acq On : 24 Sep 2021 22:03
Operator : SY/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
IBLK

Quant Time: Sep 25 01:21:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 01:18:35 2021
Response via : Initial Calibration

