

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
 Data File : VU044834.D
 Acq On : 25 Sep 2021 14:13
 Operator : SY/MD
 Sample : M3829-10DL 50X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE105D2-20210917DL

Quant Time: Sep 27 01:30:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	178582	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	314476	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	297371	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	133697	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	151035	48.459	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.920%	
35) Dibromofluoromethane	5.295	113	103486	47.969	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.940%	
50) Toluene-d8	7.902	98	403291	49.497	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.000%	
62) 4-Bromofluorobenzene	10.636	95	139142	45.978	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.960%	
Target Compounds						
					Qvalue	
44) Trichloroethene	6.549	130	54870	22.442	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
Data File : VU044834.D
Acq On : 25 Sep 2021 14:13
Operator : SY/MD
Sample : M3829-10DL 50X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE105D2-20210917DL

Quant Time: Sep 27 01:30:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

