

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
 Data File : VU044841.D
 Acq On : 25 Sep 2021 16:57
 Operator : SY/MD
 Sample : M3833-18
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE126D3-20210920

Quant Time: Sep 27 01:33:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	174379	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	309081	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	292324	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	131173	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	148879	48.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.84%	
35) Dibromofluoromethane	5.295	113	102940	48.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.10%	
50) Toluene-d8	7.903	98	397251	49.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.22%	
62) 4-Bromofluorobenzene	10.636	95	135374	45.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.02%	
Target Compounds						
						Qvalue
44) Trichloroethene	6.546	130	2713	1.13	ug/l	99
64) Tetrachloroethene	8.558	164	1583	0.80	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
Data File : VU044841.D
Acq On : 25 Sep 2021 16:57
Operator : SY/MD
Sample : M3833-18
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE126D3-20210920

Quant Time: Sep 27 01:33:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

