

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
 Data File : VU044844.D
 Acq On : 25 Sep 2021 18:07
 Operator : SY/MD
 Sample : M3829-24
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE123D2-20210918

Quant Time: Sep 27 01:33:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	174040	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	307179	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	286946	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	127502	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	150400	49.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.02%	
35) Dibromofluoromethane	5.295	113	102750	48.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.52%	
50) Toluene-d8	7.903	98	395440	49.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.38%	
62) 4-Bromofluorobenzene	10.636	95	134499	45.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.00%	
Target Compounds						
						Qvalue
44) Trichloroethene	6.546	130	4466	1.87	ug/l	98
64) Tetrachloroethene	8.556	164	2774	1.43	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
Data File : VU044844.D
Acq On : 25 Sep 2021 18:07
Operator : SY/MD
Sample : M3829-24
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE123D2-20210918

Quant Time: Sep 27 01:33:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

