

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
 Data File : VU044858.D
 Acq On : 26 Sep 2021 00:22
 Operator : SY/MD
 Sample : VU0925WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0925WBL02

Quant Time: Sep 27 02:52:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	184672	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	326106	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	306497	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	133823	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	156179	48.457	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.920%	
35) Dibromofluoromethane	5.295	113	105169	47.010	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.020%	
50) Toluene-d8	7.903	98	420785	49.802	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.600%	
62) 4-Bromofluorobenzene	10.636	95	142640	45.453	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.900%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
Data File : VU044858.D
Acq On : 26 Sep 2021 00:22
Operator : SY/MD
Sample : VU0925WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0925WBL02

Quant Time: Sep 27 02:52:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

