

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
 Data File : VU044865.D
 Acq On : 26 Sep 2021 03:07
 Operator : SY/MD
 Sample : M3859-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 1031-MW-17A

Quant Time: Sep 27 03:14:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	181870	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	320413	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	296647	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	133611	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	151850	47.840	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.680%	
35) Dibromofluoromethane	5.295	113	103485	47.080	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.160%	
50) Toluene-d8	7.902	98	405737	48.874	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.740%	
62) 4-Bromofluorobenzene	10.635	95	139511	45.245	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.500%	
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.607	62	20887	6.770	ug/l	98
16) Acetone	2.636	43	17211	9.613	ug/l	95
21) trans-1,2-Dichloroethene	3.359	96	2721	1.161	ug/l	95
27) cis-1,2-Dichloroethene	4.671	96	246862	86.572	ug/l	89
44) Trichloroethene	6.549	130	22626	9.083	ug/l	100
64) Tetrachloroethene	8.558	164	7288	3.625	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
Data File : VU044865.D
Acq On : 26 Sep 2021 03:07
Operator : SY/MD
Sample : M3859-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
1031-MW-17A

Quant Time: Sep 27 03:14:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

