

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
 Data File : VU044872.D
 Acq On : 26 Sep 2021 05:51
 Operator : SY/MD
 Sample : M3888-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 40455

Quant Time: Sep 27 03:17:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	210004	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	375287	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	344619	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	159504	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	169328	46.200	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.400%	
35) Dibromofluoromethane	5.295	113	116069	45.084	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.160%	
50) Toluene-d8	7.903	98	471305	48.471	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.940%	
62) 4-Bromofluorobenzene	10.636	95	170359	47.171	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.340%	
Target Compounds						
						Qvalue
16) Acetone	2.629	43	4180	2.022	ug/l	96
18) Methyl Acetate	2.954	43	4487	0.650	ug/l	93
95) Naphthalene	14.092	128	23060	2.792	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
Data File : VU044872.D
Acq On : 26 Sep 2021 05:51
Operator : SY/MD
Sample : M3888-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
40455

Quant Time: Sep 27 03:17:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

