

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
 Data File : VU044878.D
 Acq On : 26 Sep 2021 08:12
 Operator : SY/MD
 Sample : M3893-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-1

Quant Time: Sep 27 03:19:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	213226	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	379256	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	356114	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	164931	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	168494	45.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.56%	
35) Dibromofluoromethane	5.295	113	118094	45.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.78%	
50) Toluene-d8	7.903	98	481846	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
62) 4-Bromofluorobenzene	10.636	95	171709	47.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.10%	
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.607	62	233325	64.51	ug/l	100
6) Chloroethane	1.932	64	43432	17.65	ug/l	96
11) Tert butyl alcohol	3.179	59	13105	12.07	ug/l	99
12) 1,1-Dichloroethene	2.581	96	196796	76.65	ug/l	96
16) Acetone	2.633	43	4497	2.14	ug/l	92
24) 1,1-Dichloroethane	3.877	63	57484	10.06	ug/l	99
27) cis-1,2-Dichloroethene	4.671	96	30259	9.05	ug/l	89
39) Methylcyclohexane	6.771	83	2737	0.60	ug/l	89
40) Benzene	5.777	78	11816	0.96	ug/l	99
42) 1,2-Dichloroethane	5.800	62	28522	5.91	ug/l	99
44) Trichloroethene	6.549	130	1537	0.52	ug/l	98
49) 1,4-Dioxane	6.964	88	243529	2443.95	ug/l	91
52) Toluene	7.973	92	18630	2.54	ug/l	98
67) Ethyl Benzene	9.574	91	6287	0.48	ug/l	97
68) m/p-Xylenes	9.697	106	10382	2.06	ug/l	99
69) o-Xylene	10.105	106	4953	1.01	ug/l	100
84) 1,2,4-Trimethylbenzene	11.472	105	7240	0.74	ug/l	96
95) Naphthalene	14.089	128	26513	3.00	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
Data File : VU044878.D
Acq On : 26 Sep 2021 08:12
Operator : SY/MD
Sample : M3893-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-1

Quant Time: Sep 27 03:19:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

