

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
 Data File : VU044857.D
 Acq On : 25 Sep 2021 23:35
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 27 02:52:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.599	0.571	4.7	88	0.00
3 P	Chloromethane	0.909	0.779	14.3	86	0.00
4 C	Vinyl Chloride	0.848	0.809	4.6#	91	0.00
5 T	Bromomethane	0.450	0.438	2.7	100	0.00
6 T	Chloroethane	0.659	0.545	17.3	93	0.00
7 T	Trichlorofluoromethane	1.042	1.031	1.1	94	0.00
8 T	Diethyl Ether	0.477	0.489	-2.5	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.608	0.589	3.1	91	0.00
10 T	Methyl Iodide	0.661	0.560	15.3	75	0.00
11 T	Tert butyl alcohol	0.255	0.241	5.5	93	0.00
12 CM	1,1-Dichloroethene	0.602	0.589	2.2#	93	0.00
13 T	Acrolein	0.072	0.092	-27.8#	127	0.00
14 T	Allyl chloride	1.169	1.116	4.5	89	0.00
15 T	Acrylonitrile	0.532	0.534	-0.4	93	0.00
16 T	Acetone	0.492	0.476	3.3	95	0.00
17 T	Carbon Disulfide	1.694	1.516	10.5	86	0.00
18 T	Methyl Acetate	1.644	1.615	1.8	94	0.00
19 T	Methyl tert-butyl Ether	2.248	2.374	-5.6	97	0.00
20 T	Methylene Chloride	0.797	0.746	6.4	94	0.00
21 T	trans-1,2-Dichloroethene	0.645	0.636	1.4	92	0.00
22 T	Diisopropyl ether	2.309	2.354	-1.9	93	0.00
23 T	Vinyl Acetate	1.893	1.966	-3.9	93	0.00
24 P	1,1-Dichloroethane	1.340	1.341	-0.1	93	0.00
25 T	2-Butanone	0.747	0.741	0.8	91	0.00
26 T	2,2-Dichloropropane	1.005	0.869	13.5	82	0.00
27 T	cis-1,2-Dichloroethene	0.784	0.766	2.3	93	0.00
28 T	Bromochloromethane	0.637	0.606	4.9	90	0.00
29 T	Tetrahydrofuran	0.470	0.476	-1.3	92	0.00
30 C	Chloroform	1.347	1.318	2.2#	94	0.00
31 T	Cyclohexane	1.253	1.143	8.8	89	0.00
32 T	1,1,1-Trichloroethane	1.089	1.086	0.3	94	0.00
33 S	1,2-Dichloroethane-d4	0.873	0.893	-2.3	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.343	0.330	3.8	97	0.00
36 T	1,1-Dichloropropene	0.531	0.493	7.2	91	0.00
37 T	Ethyl Acetate	0.778	0.738	5.1	94	0.00
38 T	Carbon Tetrachloride	0.474	0.441	7.0	89	0.00
39 T	Methylcyclohexane	0.600	0.569	5.2	90	0.00
40 TM	Benzene	1.617	1.519	6.1	93	0.00
41 T	Methacrylonitrile	0.389	0.382	1.8	93	0.00
42 TM	1,2-Dichloroethane	0.636	0.609	4.2	96	0.00
43 T	Isopropyl Acetate	1.089	1.060	2.7	96	0.00
44 TM	Trichloroethene	0.389	0.351	9.8	93	0.00
45 C	1,2-Dichloropropane	0.438	0.419	4.3#	92	0.00
46 T	Dibromomethane	0.290	0.280	3.4	95	0.00
47 T	Bromodichloromethane	0.572	0.543	5.1	93	0.00
48 T	Methyl methacrylate	0.514	0.504	1.9	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
 Data File : VU044857.D
 Acq On : 25 Sep 2021 23:35
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 27 02:52:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.013	0.013	0.0	92	0.00
50 S	Toluene-d8	1.295	1.256	3.0	96	0.00
51 T	4-Methyl-2-Pentanone	0.776	0.747	3.7	93	0.00
52 CM	Toluene	0.967	0.931	3.7#	93	0.00
53 T	t-1,3-Dichloropropene	0.596	0.583	2.2	93	0.00
54 T	cis-1,3-Dichloropropene	0.646	0.626	3.1	92	0.00
55 T	1,1,2-Trichloroethane	0.413	0.391	5.3	94	0.00
56 T	Ethyl methacrylate	0.652	0.668	-2.5	96	0.00
57 T	1,3-Dichloropropane	0.736	0.704	4.3	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.145	0.256	-76.6#	168#	0.00
59 T	2-Hexanone	0.623	0.601	3.5	93	0.00
60 T	Dibromochloromethane	0.396	0.376	5.1	92	0.00
61 T	1,2-Dibromoethane	0.434	0.414	4.6	94	0.00
62 S	4-Bromofluorobenzene	0.481	0.461	4.2	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.339	0.321	5.3	90	0.00
65 PM	Chlorobenzene	1.038	1.005	3.2	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.370	0.357	3.5	93	0.00
67 C	Ethyl Benzene	1.853	1.891	-2.1#	95	0.00
68 T	m/p-Xylenes	0.709	0.713	-0.6	92	0.00
69 T	o-Xylene	0.685	0.710	-3.6	95	0.00
70 T	Styrene	1.150	1.200	-4.3	94	0.00
71 P	Bromoform	0.306	0.288	5.9	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	3.549	3.759	-5.9	93	0.00
74 T	N-amyl acetate	1.809	1.967	-8.7	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.617	1.591	1.6	95	0.00
76 T	1,2,3-Trichloropropane	1.514	1.526	-0.8	96	0.00
77 T	Bromobenzene	0.848	0.864	-1.9	93	0.00
78 T	n-propylbenzene	4.300	4.579	-6.5	93	0.00
79 T	2-Chlorotoluene	2.621	2.712	-3.5	93	0.00
80 T	1,3,5-Trimethylbenzene	3.006	3.222	-7.2	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.467	0.467	0.0	91	0.00
82 T	4-Chlorotoluene	2.984	3.163	-6.0	93	0.00
83 T	tert-Butylbenzene	2.900	3.071	-5.9	93	0.00
84 T	1,2,4-Trimethylbenzene	2.964	3.262	-10.1	93	0.00
85 T	sec-Butylbenzene	3.750	4.119	-9.8	94	0.00
86 T	p-Isopropyltoluene	3.051	3.306	-8.4	91	0.00
87 T	1,3-Dichlorobenzene	1.645	1.623	1.3	91	0.00
88 T	1,4-Dichlorobenzene	1.701	1.626	4.4	90	0.00
89 T	n-Butylbenzene	2.805	3.012	-7.4	91	0.00
90 T	Hexachloroethane	0.550	0.574	-4.4	97	0.00
91 T	1,2-Dichlorobenzene	1.744	1.686	3.3	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.392	0.424	-8.2	97	0.00
93 T	1,2,4-Trichlorobenzene	0.963	0.981	-1.9	88	0.00
94 T	Hexachlorobutadiene	0.429	0.394	8.2	83	0.00
95 T	Naphthalene	3.360	4.019	-19.6	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
Data File : VU044857.D
Acq On : 25 Sep 2021 23:35
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Sep 27 02:52:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.980	1.036	-5.7	89	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6