

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
 Data File : VU044836.D
 Acq On : 25 Sep 2021 15:00
 Operator : SY/MD
 Sample : M3784-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Sep 27 01:31:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	174774	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	313226	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	292268	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	130505	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	149727	49.086	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.180%	
35) Dibromofluoromethane	5.295	113	102704	47.796	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.600%	
50) Toluene-d8	7.902	98	398560	49.111	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.220%	
62) 4-Bromofluorobenzene	10.635	95	136661	45.338	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.680%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092521\
Data File : VU044836.D
Acq On : 25 Sep 2021 15:00
Operator : SY/MD
Sample : M3784-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Sep 27 01:31:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.378 min Scan# 1256
Delta R.T. 0.000 min
Lab File: VU044836.D
Acq: 25 Sep 2021 15:00




Instrument :
MSVOA_U
ClientSampleId :
STORAGE-BLANK-WATER-REF-5


Tgt Ion:168 Resp: 174774
Ion Ratio Lower Upper
168 100
99 66.4 51.4 77.0



#33
1,2-Dichloroethane-d4
Concen: 49.086 ug/l
RT: 5.706 min Scan# 1358
Delta R.T. -0.000 min
Lab File: VU044836.D
Acq: 25 Sep 2021 15:00

Tgt Ion: 65 Resp: 149727
Ion Ratio Lower Upper
65 100
67 51.8 0.0 104.8









#34

1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.253 min Scan# 1528
Delta R.T. -0.000 min
Lab File: VU044836.D
Acq: 25 Sep 2021 15:00

Instrument :
MSVOA_U
ClientSampleId :
STORAGE-BLANK-WATER-REF-5




Tgt Ion:114 Resp: 313226
Ion Ratio Lower Upper
114 100
63 23.5 0.0 47.8
88 16.8 0.0 32.8




#35

Dibromofluoromethane
Concen: 47.796 ug/l
RT: 5.295 min Scan# 1230
Delta R.T. -0.000 min
Lab File: VU044836.D
Acq: 25 Sep 2021 15:00

Tgt Ion:113 Resp: 102704
Ion Ratio Lower Upper
113 100
111 102.6 84.1 126.1
192 18.0 14.8 22.2









#50
Toluene-d8
Concen: 49.111 ug/l
RT: 7.902 min Scan# 2041
Delta R.T. -0.000 min
Lab File: VU044836.D
Acq: 25 Sep 2021 15:00




Instrument :
MSVOA_U
ClientSampleId :
STORAGE-BLANK-WATER-REF-5


Tgt Ion: 98 Resp: 398560
Ion Ratio Lower Upper
98 100
100 64.9 52.5 78.7



#62
4-Bromofluorobenzene
Concen: 45.338 ug/l
RT: 10.635 min Scan# 2891
Delta R.T. -0.000 min
Lab File: VU044836.D
Acq: 25 Sep 2021 15:00

Tgt Ion: 95 Resp: 136661
Ion Ratio Lower Upper
95 100
174 71.2 0.0 146.6
176 69.2 0.0 141.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.420 min Scan# 2513
Delta R.T. -0.000 min
Lab File: VU044836.D
Acq: 25 Sep 2021 15:00

Instrument :
MSVOA_U
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Tgt Ion:117 Resp: 292268
Ion Ratio Lower Upper
117 100
82 60.0 47.8 71.8
119 32.1 25.8 38.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 11.815 min Scan# 3258
Delta R.T. -0.000 min
Lab File: VU044836.D
Acq: 25 Sep 2021 15:00

Tgt Ion:152 Resp: 130505
Ion Ratio Lower Upper
152 100
115 61.5 43.5 130.7
150 157.1 0.0 350.0

