

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092523\
 Data File : VU055496.D
 Acq On : 25 Sep 2023 15:34
 Operator : MD/SY
 Sample : 04579-21
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EYJ11

Quant Time: Sep 26 01:23:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Sep 25 01:08:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	316451	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	322775	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	159999	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	96337	3.012	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	60.200%
7) Chloroethane-d5	1.911	69	96377	3.419	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	68.400%
11) 1,1-Dichloroethene-d2	2.563	65	43152	3.278	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.600%
20) 2-Butanone-d5	4.628	46	323257	55.956	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.920%
24) Chloroform-d	5.058	84	254824	4.405	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.000%
26) 1,2-Dichloroethane-d4	5.698	65	128958	4.634	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.600%
32) Benzene-d6	5.724	84	467839	4.144	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.800%
36) 1,2-Dichloropropane-d6	6.685	67	155251	4.602	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.000%
41) Toluene-d8	7.894	98	387812	3.769	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.400%
43) trans-1,3-Dichloroprop...	8.177	79	52210	4.410	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.200%
46) 2-Hexanone-d5	8.631	63	181006	50.269	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.540%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	125415	4.961	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.200%
66) 1,2-Dichlorobenzene-d4	12.190	152	149003	4.614	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092523\
Data File : VU055496.D
Acq On : 25 Sep 2023 15:34
Operator : MD/SY
Sample : 04579-21
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EYJ11

Quant Time: Sep 26 01:23:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Sep 25 01:08:17 2023
Response via : Initial Calibration

