

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092525\
 Data File : VU063641.D
 Acq On : 25 Sep 2025 09:24
 Operator : MD/SY
 Sample : VSTDCCC0.5
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC0.5

Quant Time: Sep 26 05:10:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM092425.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Sep 25 04:13:39 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
2 S	Vinyl Chloride-d3	0.336	0.335	0.3	106	0.00
3 T	Vinyl chloride	0.660	0.637	3.5	101	0.00
4 S	1,2-Dichloroethane-d4	0.199	0.184	7.5	109	0.00
5 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
6 T	Trichloroethene	0.288	0.260	9.7	100	0.00
7 S	1,2-Dichloropropane-d6	0.131	0.135	-3.1	107	0.00
8 S	Toluene-d8	0.420	0.386	8.1	104	0.00
9 T	1,2-Dibromoethane	0.159	0.162	-1.9	101	0.00
10 S	1,1,2,2-Tetrachloroethane-d	0.117	0.113	3.4	103	0.00
11 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.02
12	1,2,3-Trichloropropane	0.441	0.464	-5.2	98	0.00
13 T	1,2-Dibromo-3-chloropropane	0.062	0.063	-1.6	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0