

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092525\
 Data File : VU063644.D
 Acq On : 25 Sep 2025 13:00
 Operator : MD/SY
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC0.5EC

Quant Time: Sep 26 05:11:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM092425.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Sep 25 04:13:39 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
2 S	Vinyl Chloride-d3	0.336	0.332	1.2	93	0.00
3 T	Vinyl chloride	0.660	0.615	6.8	86	0.00
4 S	1,2-Dichloroethane-d4	0.199	0.203	-2.0	106	0.00
5 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
6 T	Trichloroethene	0.288	0.277	3.8	94	0.00
7 S	1,2-Dichloropropane-d6	0.131	0.140	-6.9	98	0.00
8 S	Toluene-d8	0.420	0.411	2.1	98	0.00
9 T	1,2-Dibromoethane	0.159	0.166	-4.4	92	0.00
10 S	1,1,2,2-Tetrachloroethane-d	0.117	0.125	-6.8	101	0.00
11 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.02
12	1,2,3-Trichloropropane	0.441	0.493	-11.8	94	0.00
13 T	1,2-Dibromo-3-chloropropane	0.062	0.068	-9.7	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0