

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU092618\
 Data File : VU027203.D
 Acq On : 27 Sep 2018 01:57
 Operator : MD/SY
 Sample : J5048-06
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-80-SEP18

Quant Time: Sep 27 02:48:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	112489	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	187835	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	182610	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	89936	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	107418	47.45	ug/l	0.00
Spiked Amount			Recovery	=	94.90%	
35) Dibromofluoromethane	4.88	113	76869	47.21	ug/l	0.00
Spiked Amount			Recovery	=	94.42%	
50) Toluene-d8	7.57	98	295534	47.63	ug/l	0.00
Spiked Amount			Recovery	=	95.26%	
62) 4-Bromofluorobenzene	10.31	95	100402	47.16	ug/l	0.00
Spiked Amount			Recovery	=	94.32%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	128757	54.473	ug/l	100
6) Chloroethane	1.58	64	1287	0.852	ug/l #	43
11) Tert butyl alcohol	2.84	59	963	1.856	ug/l #	72
12) 1,1-Dichloroethene	2.29	96	4058	2.612	ug/l	97
13) Acrolein	2.20	56	526	1.880	ug/l #	23
16) Acetone	2.32	43	3152	Below Cal		94
21) trans-1,2-Dichloroethene	2.99	96	813	0.490	ug/l #	72
24) 1,1-Dichloroethane	3.45	63	4504	1.213	ug/l #	94
27) cis-1,2-Dichloroethene	4.23	96	328649	175.439	ug/l	89
31) Cyclohexane	4.99	56	2809	0.260	ug/l #	1
32) 1,1,1-Trichloroethane	4.92	97	762	0.277	ug/l #	53
36) 1,1-Dichloropropene	4.98	75	10051	4.009	ug/l #	50
38) Carbon Tetrachloride	4.99	117	11754	4.927	ug/l #	18
44) Trichloroethene	6.19	130	40565	25.003	ug/l	97
48) Methyl methacrylate	6.65	41	22	1.020	ug/l #	17
55) 1,1,2-Trichloroethane	7.94	97	507	0.274	ug/l #	13
60) Dibromochloromethane	8.23	129	94657	52.136	ug/l #	11
64) Tetrachloroethene	8.23	164	99975	71.048	ug/l	95
68) m/p-Xylenes	9.39	106	67	1.521	ug/l #	54
76) 1,2,3-Trichloropropane	10.31	75	52783	19.319	ug/l #	36
81) trans-1,4-Dichloro-2-buten	10.31	75	52783	61.823	ug/l #	100
89) n-Butylbenzene	11.89	91	401	2.454	ug/l #	70
95) Naphthalene	13.75	128	800	1.793	ug/l #	69


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU092618\
Data File : VU027203.D
Acq On : 27 Sep 2018 01:57
Operator : MD/SY
Sample : J5048-06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-80-SEP18

Quant Time: Sep 27 02:48:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.99 min Scan# 1186

Delta R.T. -0.00 min

Lab File: VU027203.D

Acq: 27 Sep 2018 01:57

Instrument :

MSVOA_U

ClientSampleId :


MW-80-SEP18

Tgt Ion:168 Resp: 112489

Ion Ratio Lower Upper

168 100

99 61.2 48.1 72.1



Abundance Ion 167.90 (167.60 to 168.60): VU026969.D (-1169) (-)

Ion 98.90 (98.60 to 99.60): VU026969.D (-1169) (-)

4.99

50000

40000

30000



20000

10000

0

4.90 4.95 5.00 5.05 5.10

Time-->



#4

Vinyl Chloride

Concen: 54.473 ug/l

RT: 1.40 min Scan# 71

Delta R.T. -0.00 min

Lab File: VU027203.D


Acq: 27 Sep 2018 01:57

Tgt Ion: 62 Resp: 128757

Ion Ratio Lower Upper

62 100

64 31.6 25.5 38.3



Abundance Ion 61.90 (61.60 to 62.60): VU026969.D (-63) (-)

Ion 63.90 (63.60 to 64.60): VU026969.D (-63) (-)

1.40

150000


100000


50000

0

1.35 1.40 1.45 1.50

Time-->





#6

Chloroethane

Concen: 0.852 ug/l

RT: 1.58 min Scan# 127

Delta R.T. -0.11 min

Lab File: VU027203.D

Acq: 27 Sep 2018 01:57

Instrument :

MSVOA_U

ClientSampleId :


MW-80-SEP18

Tgt Ion: 64 Resp: 1287

Ion Ratio Lower Upper

64 100

66 0.0 25.3 37.9#



Abundance Ion 63.90 (63.60 to 64.60): VU0

Ion 65.90 (65.60 to 66.60): VU0

1.58

2500

2000



1500

1000

500

0

Time-->



#11

Tert butyl alcohol

Concen: 1.856 ug/l

RT: 2.84 min Scan# 517

Delta R.T. 0.01 min

Lab File: VU027203.D


Acq: 27 Sep 2018 01:57

Tgt Ion: 59 Resp: 963

Ion Ratio Lower Upper

59 100

57 0.0 8.5 12.7#



Abundance Ion 59.00 (58.70 to 59.70): VU0

Ion 57.00 (56.70 to 57.70): VU0

2.84

500

400


300


200

100

0

Time-->









#12
 1,1-Dichloroethene
 Concen: 2.612 ug/l
 RT: 2.29 min Scan# 346
 Delta R.T. 0.00 min
 Lab File: VU027203.D
 Acq: 27 Sep 2018 01:57




Instrument :
 MSVOA_U
 ClientSampleId :
 MW-80-SEP18


Tgt Ion: 96 Resp: 4058
 Ion Ratio Lower Upper
 96 100
 61 190.5 156.1 234.1
 98 65.7 51.4 77.0



#13
 Acrolein
 Concen: 1.880 ug/l
 RT: 2.20 min Scan# 319
 Delta R.T. 0.00 min
 Lab File: VU027203.D
 Acq: 27 Sep 2018 01:57

Tgt Ion: 56 Resp: 526
 Ion Ratio Lower Upper
 56 100
 55 7.4 57.1 85.7#





#16

Acetone

Concen: Below Cal

RT: 2.32 min Scan# 358

Delta R.T. 0.00 min

Lab File: VU027203.D

Acq: 27 Sep 2018 01:57

Instrument :

MSVOA_U

ClientSampleId :


MW-80-SEP18

Tgt Ion: 43 Resp: 3152

Ion Ratio Lower Upper




43 100

58 34.0 24.6 36.8



Abundance Ion 43.00 (42.70 to 43.70): VU026969.D (-335) (-)

Ion 58.00 (57.70 to 58.70): VU026969.D (-335) (-)



#21

trans-1,2-Dichloroethene

Concen: 0.490 ug/l

RT: 2.99 min Scan# 565

Delta R.T. 0.01 min

Lab File: VU027203.D

Acq: 27 Sep 2018 01:57


Tgt Ion: 96 Resp: 813

Ion Ratio Lower Upper

96 100

61 127.6 131.8 197.6#



98 40.9 50.9 76.3#




Abundance Ion 95.90 (95.60 to 96.60): VU026969.D (-549) (-)

Ion 60.90 (60.60 to 61.60): VU026969.D (-549) (-)

Ion 97.90 (97.60 to 98.60): VU026969.D (-549) (-)









#24
 1,1-Dichloroethane
 Concen: 1.213 ug/l
 RT: 3.45 min Scan# 707
 Delta R.T. -0.00 min
 Lab File: VU027203.D
 Acq: 27 Sep 2018 01:57




Instrument :
 MSVOA_U
 ClientSampleId :
 MW-80-SEP18


Tgt Ion: 63 Resp: 4504
 Ion Ratio Lower Upper
 63 100
 98 5.5 3.3 9.8
 100 0.0 2.0 5.8#



#27
 cis-1,2-Dichloroethene
 Concen: 175.439 ug/l
 RT: 4.23 min Scan# 950
 Delta R.T. -0.00 min
 Lab File: VU027203.D
 Acq: 27 Sep 2018 01:57

Tgt Ion: 96 Resp: 328649
 Ion Ratio Lower Upper
 96 100
 61 149.1 0.0 337.2
 98 64.2 0.0 127.6





#31

Cyclohexane

Concen: 0.260 ug/l

RT: 4.99 min Scan# 1186


Delta R.T. -0.01 min

Lab File: VU027203.D




Acq: 27 Sep 2018 01:57

Instrument :
MSVOA_U
ClientSampleId :
MW-80-SEP18

Tgt Ion:	56	Resp:	2809
Ion	Ratio	Lower	Upper
56	100		
69	160.2	23.0	34.4#
84	128.1	61.4	92.2#



Abundance Ion 56.00 (55.70 to 56.70): VU0
Ion 69.00 (68.70 to 69.70): VU0
Ion 84.00 (83.70 to 84.70): VU0



#32

1,1,1-Trichloroethane

Concen: 0.277 ug/l


RT: 4.92 min Scan# 1164

Delta R.T. -0.00 min



Lab File: VU027203.D


Acq: 27 Sep 2018 01:57

Tgt Ion:	97	Resp:	762
Ion	Ratio	Lower	Upper
97	100		
99	0.0	51.8	77.6#
61	53.8	41.1	61.7



Abundance Ion 96.90 (96.60 to 97.60): VU0
Ion 98.90 (98.60 to 99.60): VU0
Ion 60.90 (60.60 to 61.60): VU0









#33
 1,2-Dichloroethane-d4
 Concen: 47.446 ug/l
 RT: 5.31 min Scan# 1287
 Delta R.T. -0.00 min
 Lab File: VU027203.D
 Acq: 27 Sep 2018 01:57




Instrument :
 MSVOA_U
 ClientSampleId :
 MW-80-SEP18


Tgt Ion: 65 Resp: 107418
 Ion Ratio Lower Upper
 65 100
 67 52.9 0.0 105.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 5.89 min Scan# 1466
 Delta R.T. -0.00 min
 Lab File: VU027203.D
 Acq: 27 Sep 2018 01:57

Tgt Ion: 114 Resp: 187835
 Ion Ratio Lower Upper
 114 100
 63 24.1 0.0 49.2
 88 17.3 0.0 33.8









#35
 Dibromofluoromethane
 Concen: 47.212 ug/l
 RT: 4.88 min Scan# 1154
 Delta R.T. 0.00 min
 Lab File: VU027203.D
 Acq: 27 Sep 2018 01:57




Instrument :
 MSVOA_U
 ClientSampleId :
 MW-80-SEP18


Tgt Ion:113 Resp: 76869
 Ion Ratio Lower Upper
 113 100
 111 104.3 82.2 123.4
 192 19.4 15.2 22.8



#36
 1,1-Dichloropropene
 Concen: 4.009 ug/l
 RT: 4.98 min Scan# 1185
 Delta R.T. -0.15 min
 Lab File: VU027203.D
 Acq: 27 Sep 2018 01:57

Tgt Ion: 75 Resp: 10051
 Ion Ratio Lower Upper
 75 100
 110 7.9 16.1 48.2#
 77 0.0 24.6 36.8#





#38

Carbon Tetrachloride

Concen: 4.927 ug/l

RT: 4.99 min Scan# 1186

Delta R.T. -0.15 min

Lab File: VU027203.D

Acq: 27 Sep 2018 01:57

Instrument :

MSVOA_U

ClientSampleId :

MW-80-SEP18


Tgt Ion:117 Resp: 11754

Ion Ratio Lower Upper

117 100

119 5.7 74.4 111.6#




121 0.0 24.2 36.4#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#44

Trichloroethene

Concen: 25.003 ug/l

RT: 6.19 min Scan# 1560

Delta R.T. 0.00 min

Lab File: VU027203.D


Acq: 27 Sep 2018 01:57

Tgt Ion:130 Resp: 40565

Ion Ratio Lower Upper



130 100


95 100.9 0.0 208.8



Abundance Ion 129.80 (129.50 to 130.50): V

Ion 94.90 (94.60 to 95.60): VU0









#48
Methyl methacrylate
Concen: 1.020 ug/l
RT: 6.65 min Scan# 1702
Delta R.T. 0.02 min
Lab File: VU027203.D
Acq: 27 Sep 2018 01:57




Instrument :
MSVOA_U
ClientSampleId :
MW-80-SEP18


Tgt Ion: 41 Resp: 22
Ion Ratio Lower Upper
41 100
69 0.0 60.6 90.8#
39 0.0 42.8 64.2#



#50
Toluene-d8
Concen: 47.629 ug/l
RT: 7.57 min Scan# 1988
Delta R.T. -0.00 min
Lab File: VU027203.D
Acq: 27 Sep 2018 01:57

Tgt Ion: 98 Resp: 295534
Ion Ratio Lower Upper
98 100
100 64.9 52.1 78.1









#55
 1,1,2-Trichloroethane
 Concen: 0.274 ug/l
 RT: 7.94 min Scan# 2106
 Delta R.T. -0.13 min
 Lab File: VU027203.D
 Acq: 27 Sep 2018 01:57




Instrument :
 MSVOA_U
 ClientSampleId :
 MW-80-SEP18


Tgt Ion: 97 Resp: 507
 Ion Ratio Lower Upper
 97 100
 83 0.0 73.1 109.7#
 85 0.0 46.7 70.1#
 99 0.0 50.0 75.0#



#60
 Dibromochloromethane
 Concen: 52.136 ug/l
 RT: 8.23 min Scan# 2194
 Delta R.T. -0.25 min
 Lab File: VU027203.D
 Acq: 27 Sep 2018 01:57

Tgt Ion: 129 Resp: 94657
 Ion Ratio Lower Upper
 129 100
 127 0.0 38.1 114.5#





#62

4-Bromofluorobenzene

Concen: 47.157 ug/l

RT: 10.31 min Scan# 2841

Delta R.T. -0.00 min

Lab File: VU027203.D

Acq: 27 Sep 2018 01:57

Instrument :

MSVOA_U

ClientSampleId :

MW-80-SEP18


Tgt Ion: 95 Resp: 100402

Ion Ratio Lower Upper

95 100

174 76.0 0.0 150.2

176 71.8 0.0 145.2




Abundance Ion 94.90 (94.60 to 95.60): VU026969.D (-2828) (-)

Ion 173.80 (173.50 to 174.50): VU026969.D (-2828) (-)

Ion 175.80 (175.50 to 176.50): VU026969.D (-2828) (-)

Time-->




Abundance Ion 94.90 (94.60 to 95.60): VU027203.D (-2748) (-)

Ion 173.80 (173.50 to 174.50): VU027203.D (-2748) (-)

Ion 175.80 (175.50 to 176.50): VU027203.D (-2748) (-)

Time-->



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 9.09 min Scan# 2463

Delta R.T. 0.00 min

Lab File: VU027203.D

Acq: 27 Sep 2018 01:57


Tgt Ion: 117 Resp: 182610

Ion Ratio Lower Upper

117 100

82 60.5 48.0 72.0

119 31.8 25.8 38.8




Abundance Ion 116.90 (116.60 to 117.60): VU026969.D (-2449) (-)

Ion 82.00 (81.70 to 82.70): VU026969.D (-2449) (-)

Ion 118.90 (118.60 to 119.60): VU026969.D (-2449) (-)

Time-->




Abundance Ion 116.90 (116.60 to 117.60): VU027203.D (-2370) (-)

Ion 82.00 (81.70 to 82.70): VU027203.D (-2370) (-)

Ion 118.90 (118.60 to 119.60): VU027203.D (-2370) (-)


Time-->






#64
Tetrachloroethene
Concen: 71.048 ug/l
RT: 8.23 min Scan# 2194
Delta R.T. 0.00 min
Lab File: VU027203.D
Acq: 27 Sep 2018 01:57

Instrument :
MSVOA_U
ClientSampleId :
MW-80-SEP18

Tgt Ion:164 Resp: 99975
Ion Ratio Lower Upper
164 100
166 123.7 104.4 156.6
129 94.2 78.7 118.1
131 89.4 73.9 110.9




Abundance Ion 163.80 (163.50 to 164.50): V
Ion 165.80 (165.50 to 166.50): V
Ion 128.80 (128.50 to 129.50): V
Ion 130.80 (130.50 to 131.50): V






#68
m/p-Xylenes
Concen: 1.521 ug/l
RT: 9.39 min Scan# 2554
Delta R.T. 0.01 min
Lab File: VU027203.D
Acq: 27 Sep 2018 01:57

Tgt Ion:106 Resp: 67
Ion Ratio Lower Upper
106 100
91 279.1 166.5 249.7#



Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): V






#72




1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 11.48 min Scan# 3207
Delta R.T. -0.00 min
Lab File: VU027203.D
Acq: 27 Sep 2018 01:57

Instrument :
MSVOA_U
ClientSampleId :
MW-80-SEP18

Tgt Ion:152 Resp: 89936
Ion Ratio Lower Upper
152 100
115 59.7 44.9 134.5
150 156.9 0.0 355.8




Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#76


1,2,3-Trichloropropane
Concen: 19.319 ug/l
RT: 10.31 min Scan# 2841
Delta R.T. -0.19 min
Lab File: VU027203.D
Acq: 27 Sep 2018 01:57

Tgt Ion: 75 Resp: 52783
Ion Ratio Lower Upper
75 100
77 1.2 20.9 62.8#



Abundance Ion 74.90 (74.60 to 75.60): VU0
Ion 77.00 (76.70 to 77.70): VU0









#81
 trans-1,4-Dichloro-2-butene
 Concen: 61.823 ug/l
 RT: 10.31 min Scan# 2841
 Delta R.T. -0.21 min
 Lab File: VU027203.D
 Acq: 27 Sep 2018 01:57




Instrument :
 MSVOA_U
 ClientSampleId :
 MW-80-SEP18


Tgt Ion: 75 Resp: 52783
 Ion Ratio Lower Upper
 75 100
 53 0.1 0.0 0.0#
 89 0.0 0.0 0.0



#89
 n-Butylbenzene
 Concen: 2.454 ug/l
 RT: 11.89 min Scan# 3333
 Delta R.T. 0.00 min
 Lab File: VU027203.D
 Acq: 27 Sep 2018 01:57

Tgt Ion: 91 Resp: 401
 Ion Ratio Lower Upper
 91 100
 92 37.7 27.1 81.2
 134 0.0 11.7 35.0#






#95
Naphthalene
Concen: 1.793 ug/l
RT: 13.75 min Scan# 3911
Delta R.T. 0.01 min
Lab File: VU027203.D
Acq: 27 Sep 2018 01:57

Instrument :
MSVOA_U
ClientSampleId :
MW-80-SEP18

Tgt Ion:128 Resp: 800
Ion Ratio Lower Upper
128 100
127 0.0 10.6 15.8#
129 0.0 8.6 12.8#



Abundance Ion 128.00 (127.70 to 128.70): V
Ion 127.00 (126.70 to 127.70): V
Ion 129.00 (128.70 to 129.70): V

