

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092618\
 Data File : VU027140.D
 Acq On : 25 Sep 2018 21:03
 Operator : MD/SY
 Sample : VU0926WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0926WBL01

Quant Time: Sep 26 06:46:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	92324	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	155827	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	147505	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	67474	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	96742	52.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.12%	
35) Dibromofluoromethane	4.88	113	66800	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
50) Toluene-d8	7.57	98	251857	48.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.86%	
62) 4-Bromofluorobenzene	10.31	95	79293	44.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.78%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092618\
Data File : VU027140.D
Acq On : 25 Sep 2018 21:03
Operator : MD/SY
Sample : VU0926WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0926WBL01

Quant Time: Sep 26 06:46:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

