

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092618\
 Data File : VU027142.D
 Acq On : 25 Sep 2018 21:50
 Operator : MD/SY
 Sample : J4968-01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB03-20180916

Quant Time: Sep 26 06:54:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	95019	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	159439	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	151557	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	68663	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	99386	51.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.94%	
35) Dibromofluoromethane	4.89	113	68678	49.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.38%	
50) Toluene-d8	7.57	98	249308	47.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.66%	
62) 4-Bromofluorobenzene	10.31	95	81134	44.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.78%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092618\
Data File : VU027142.D
Acq On : 25 Sep 2018 21:50
Operator : MD/SY
Sample : J4968-01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB03-20180916

Quant Time: Sep 26 06:54:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

