

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092618\
 Data File : VU027153.D
 Acq On : 26 Sep 2018 02:06
 Operator : MD/SY
 Sample : J4968-11
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SA-PZ143I-20180916

Quant Time: Sep 26 07:18:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	90720	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	153368	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	146749	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	68861	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	98101	53.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.46%	
35) Dibromofluoromethane	4.89	113	66326	49.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.78%	
50) Toluene-d8	7.57	98	249497	49.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.50%	
62) 4-Bromofluorobenzene	10.31	95	80294	46.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.38%	
Target Compounds						
6) Chloroethane	1.70	64	13540	11.117	ug/l	100
12) 1,1-Dichloroethene	2.28	96	7727	6.167	ug/l	93
24) 1,1-Dichloroethane	3.45	63	279541	93.374	ug/l	99
32) 1,1,1-Trichloroethane	4.92	97	16506	7.439	ug/l	98
85) sec-Butylbenzene	11.33	105	2449	0.469	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092618\
Data File : VU027153.D
Acq On : 26 Sep 2018 02:06
Operator : MD/SY
Sample : J4968-11
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SA-PZ143I-20180916

Quant Time: Sep 26 07:18:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

