

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092618\
 Data File : VU027155.D
 Acq On : 26 Sep 2018 02:52
 Operator : MD/SY
 Sample : J4980-09
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BPSI-TT-MW304D-20180917

Quant Time: Sep 26 07:21:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	89834	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	153219	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	147215	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	66467	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	97200	53.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.52%	
35) Dibromofluoromethane	4.89	113	67111	50.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.06%	
50) Toluene-d8	7.57	98	246392	48.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.36%	
62) 4-Bromofluorobenzene	10.31	95	77859	44.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.66%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092618\
Data File : VU027155.D
Acq On : 26 Sep 2018 02:52
Operator : MD/SY
Sample : J4980-09
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BPSI-TT-MW304D-20180917

Quant Time: Sep 26 07:21:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

