

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092618\
 Data File : VU027164.D
 Acq On : 26 Sep 2018 09:49
 Operator : MD/SY
 Sample : VU0926WBL02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0926WBL02

Quant Time: Sep 27 01:26:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	99803	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	168055	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	157345	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	69268	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	101344	50.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.90%	
35) Dibromofluoromethane	4.88	113	70563	48.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.88%	
50) Toluene-d8	7.57	98	266887	48.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.14%	
62) 4-Bromofluorobenzene	10.31	95	85236	44.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.50%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092618\
Data File : VU027164.D
Acq On : 26 Sep 2018 09:49
Operator : MD/SY
Sample : VU0926WBL02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0926WBL02

Quant Time: Sep 27 01:26:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

