

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092618\
 Data File : VU027166.D
 Acq On : 26 Sep 2018 10:47
 Operator : MD/SY
 Sample : J5032-09
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-16-092018

Quant Time: Sep 27 01:35:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	98494	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	166887	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	157463	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	71541	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	99950	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
35) Dibromofluoromethane	4.88	113	69227	47.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.72%	
50) Toluene-d8	7.57	98	263574	47.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.62%	
62) 4-Bromofluorobenzene	10.31	95	84927	44.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.80%	
Target Compounds						
4) Vinyl Chloride	1.40	62	21781	10.524	ug/l	98
27) cis-1,2-Dichloroethene	4.23	96	174104	106.146	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092618\
Data File : VU027166.D
Acq On : 26 Sep 2018 10:47
Operator : MD/SY
Sample : J5032-09
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-16-092018

Quant Time: Sep 27 01:35:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

