

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092618\
 Data File : VU027169.D
 Acq On : 26 Sep 2018 11:56
 Operator : MD/SY
 Sample : J5032-06
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-18-092018

Quant Time: Sep 27 02:00:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	90138	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	153869	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	146580	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	66361	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	95914	52.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.74%	
35) Dibromofluoromethane	4.88	113	65364	49.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.02%	
50) Toluene-d8	7.57	98	247221	48.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.28%	
62) 4-Bromofluorobenzene	10.31	95	79925	45.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.66%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092618\
Data File : VU027169.D
Acq On : 26 Sep 2018 11:56
Operator : MD/SY
Sample : J5032-06
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-18-092018

Quant Time: Sep 27 02:00:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

