

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092618\
 Data File : VU027172.D
 Acq On : 26 Sep 2018 13:06
 Operator : MD/SY
 Sample : J5032-08
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-17-092018

Quant Time: Sep 27 02:05:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	92253	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	161042	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	151431	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	69685	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	99770	53.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.46%	
35) Dibromofluoromethane	4.88	113	68060	48.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.52%	
50) Toluene-d8	7.57	98	253711	47.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.38%	
62) 4-Bromofluorobenzene	10.31	95	82229	45.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.10%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092618\
Data File : VU027172.D
Acq On : 26 Sep 2018 13:06
Operator : MD/SY
Sample : J5032-08
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-17-092018

Quant Time: Sep 27 02:05:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

