

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092618\
 Data File : VU027190.D
 Acq On : 26 Sep 2018 20:54
 Operator : MD/SY
 Sample : VU0926WBL03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0926WBL03

Quant Time: Sep 27 02:51:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	92883	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	158201	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	151868	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	70280	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	97777	52.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.60%	
35) Dibromofluoromethane	4.88	113	67250	49.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.08%	
50) Toluene-d8	7.57	98	256200	49.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.04%	
62) 4-Bromofluorobenzene	10.31	95	82750	46.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.30%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092618\
Data File : VU027190.D
Acq On : 26 Sep 2018 20:54
Operator : MD/SY
Sample : VU0926WBL03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0926WBL03

Quant Time: Sep 27 02:51:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

