

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092618\
 Data File : VU027192.D
 Acq On : 26 Sep 2018 21:40
 Operator : MD/SY
 Sample : J5048-19
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB-2-SEP18

Quant Time: Sep 27 02:57:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	92840	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	157296	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	151934	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	69430	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	97826	52.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.70%	
35) Dibromofluoromethane	4.88	113	67234	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
50) Toluene-d8	7.57	98	253933	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
62) 4-Bromofluorobenzene	10.31	95	84311	47.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.58%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092618\
Data File : VU027192.D
Acq On : 26 Sep 2018 21:40
Operator : MD/SY
Sample : J5048-19
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB-2-SEP18

Quant Time: Sep 27 02:57:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

