

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092618\
 Data File : VU027195.D
 Acq On : 26 Sep 2018 22:50
 Operator : MD/SY
 Sample : J5049-02DL 4X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-95-SEP18DL

Quant Time: Sep 27 04:08:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	94766	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	159527	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	151345	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	70391	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	97625	51.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.36%	
35) Dibromofluoromethane	4.88	113	69461	50.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.46%	
50) Toluene-d8	7.57	98	251790	47.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.56%	
62) 4-Bromofluorobenzene	10.31	95	82112	45.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.82%	
Target Compounds						
4) Vinyl Chloride	1.40	62	17617	8.847	ug/l	99
27) cis-1,2-Dichloroethene	4.23	96	56200	35.611	ug/l	91
65) Chlorobenzene	9.12	112	22803	5.682	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092618\
Data File : VU027195.D
Acq On : 26 Sep 2018 22:50
Operator : MD/SY
Sample : J5049-02DL 4X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-95-SEP18DL

Quant Time: Sep 27 04:08:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

