

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092618\
 Data File : VU027198.D
 Acq On : 27 Sep 2018 00:00
 Operator : MD/SY
 Sample : J5048-01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CMT-18-2-SEP18

Quant Time: Sep 27 04:06:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	90249	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	150661	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	144332	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	68930	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	92453	50.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.80%	
35) Dibromofluoromethane	4.88	113	63521	48.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.28%	
50) Toluene-d8	7.57	98	240516	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
62) 4-Bromofluorobenzene	10.31	95	79628	46.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.26%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	166831	87.975	ug/l	99
12) 1,1-Dichloroethene	2.29	96	19020	15.259	ug/l	92
21) trans-1,2-Dichloroethene	2.99	96	40122	30.148	ug/l	99
27) cis-1,2-Dichloroethene	4.23	96	6376255	4242.556	ug/l	93
44) Trichloroethene	6.19	130	135026	103.761	ug/l	95
52) Toluene	7.64	92	2762	0.788	ug/l	94
59) 2-Hexanone	8.38	43	3293	1.576	ug/l	91
95) Naphthalene	13.75	128	275	1.736	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092618\
Data File : VU027198.D
Acq On : 27 Sep 2018 00:00
Operator : MD/SY
Sample : J5048-01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CMT-18-2-SEP18

Quant Time: Sep 27 04:06:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

