

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092618\
 Data File : VU027202.D
 Acq On : 27 Sep 2018 01:34
 Operator : MD/SY
 Sample : J5048-05
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DMW-35A-SEP18

Quant Time: Sep 27 03:46:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	115500	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	191625	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	184765	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	93890	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	109334	47.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.06%	
35) Dibromofluoromethane	4.88	113	78133	47.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.08%	
50) Toluene-d8	7.57	98	291503	46.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.10%	
62) 4-Bromofluorobenzene	10.31	95	104394	48.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.12%	
Target Compounds						
16) Acetone	2.32	43	168185	128.114	ug/l	Qvalue 100
27) cis-1,2-Dichloroethene	4.23	96	18759	9.753	ug/l	91
44) Trichloroethene	6.19	130	43239	26.124	ug/l	99
64) Tetrachloroethene	8.23	164	46130	32.400	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092618\
Data File : VU027202.D
Acq On : 27 Sep 2018 01:34
Operator : MD/SY
Sample : J5048-05
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DMW-35A-SEP18

Quant Time: Sep 27 03:46:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

