

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092618\
 Data File : VU027204.D
 Acq On : 27 Sep 2018 02:21
 Operator : MD/SY
 Sample : J5048-07
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-83-SEP18

Quant Time: Sep 27 03:40:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	112615	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	189082	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	181182	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	87010	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	107898	47.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.20%	
35) Dibromofluoromethane	4.88	113	78569	47.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.88%	
50) Toluene-d8	7.57	98	295107	47.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.50%	
62) 4-Bromofluorobenzene	10.31	95	99891	46.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.22%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	18973	8.018	ug/l	98
12) 1,1-Dichloroethene	2.28	96	1148	0.738	ug/l	94
25) 2-Butanone	4.28	43	7149	3.753	ug/l	93
27) cis-1,2-Dichloroethene	4.23	96	66559	35.491	ug/l	90
44) Trichloroethene	6.19	130	35575	21.783	ug/l	95
64) Tetrachloroethene	8.23	164	16107	11.537	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092618\
Data File : VU027204.D
Acq On : 27 Sep 2018 02:21
Operator : MD/SY
Sample : J5048-07
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-83-SEP18

Quant Time: Sep 27 03:40:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

