

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092618\
 Data File : VU027211.D
 Acq On : 27 Sep 2018 05:04
 Operator : MD/SY
 Sample : J5048-09
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CMT-22-6-SEP18

Quant Time: Sep 27 08:19:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	107278	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	181069	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	173183	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	82095	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	106591	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
35) Dibromofluoromethane	4.88	113	75285	47.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.94%	
50) Toluene-d8	7.57	98	285317	47.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.40%	
62) 4-Bromofluorobenzene	10.31	95	95754	46.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.30%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.23	96	12053	6.747	ug/l	89
44) Trichloroethene	6.19	130	6440	4.118	ug/l	96
64) Tetrachloroethene	8.23	164	1158	0.868	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092618\
Data File : VU027211.D
Acq On : 27 Sep 2018 05:04
Operator : MD/SY
Sample : J5048-09
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CMT-22-6-SEP18

Quant Time: Sep 27 08:19:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

