

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092620\
 Data File : VU040301.D
 Acq On : 25 Sep 2020 15:38
 Operator : SY/MD
 Sample : L4147-10
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFSJ0

Quant Time: Sep 26 06:40:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 26 04:43:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	126260	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	125250	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	58476	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	31767	4.61	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.20%
7) Chloroethane-d5	1.92	69	30530	5.32	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.40%
11) 1,1-Dichloroethene-d2	2.58	63	53455	3.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.40%
20) 2-Butanone-d5	4.66	46	157294	57.05	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.10%
24) Chloroform-d	5.08	84	73928	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.80%
26) 1,2-Dichloroethane-d4	5.72	65	45800	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.20%
32) Benzene-d6	5.74	84	133636	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
36) 1,2-Dichloropropane-d6	6.70	67	46675	4.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.80%
41) Toluene-d8	7.91	98	110234	4.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.40%
43) trans-1,3-Dichloropropene-	8.19	79	17681	3.99	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.80%
46) 2-Hexanone-d5	8.64	63	108018	51.50	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.00%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	40258	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	47228	4.99	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.80%

Target Compounds					Ovalue
3) Chloromethane	1.53	50	12468	1.039 ug/L	97
13) Acetone	2.67	43	17852	6.994 ug/L	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092620\
Data File : VU040301.D
Acq On : 25 Sep 2020 15:38
Operator : SY/MD
Sample : L4147-10
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFSJ0

Quant Time: Sep 26 06:40:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 26 04:43:23 2020
Response via : Initial Calibration

