

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092623\
 Data File : VU055512.D
 Acq On : 26 Sep 2023 13:07
 Operator : MD/SY
 Sample : VIBLK142
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK142

Quant Time: Sep 27 04:48:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 27 04:44:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	309601	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	312476	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	152737	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	126186	4.033	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	80.600%	
7) Chloroethane-d5	1.911	69	95743	3.472	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	69.400%	
11) 1,1-Dichloroethene-d2	2.563	65	49210	3.821	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	76.400%	
20) 2-Butanone-d5	4.634	46	225497	39.898	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	79.800%	
24) Chloroform-d	5.058	84	217469	3.842	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	76.800%	
26) 1,2-Dichloroethane-d4	5.701	65	105130	3.862	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	77.200%	
32) Benzene-d6	5.724	84	444994	4.071	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	81.400%	
36) 1,2-Dichloropropane-d6	6.688	67	135795	4.158	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	83.200%	
41) Toluene-d8	7.897	98	385429	3.869	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	77.400%	
43) trans-1,3-Dichloroprop...	8.180	79	47797	4.171	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	83.400%	
46) 2-Hexanone-d5	8.634	63	134940	38.711	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	77.420%	
56) 1,1,2,2-Tetrachloroeth...	10.756	84	97243	3.973	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	79.400%	
66) 1,2-Dichlorobenzene-d4	12.193	152	130825	4.244	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	84.800%	
Target Compounds						Qvalue
16) Methylene chloride	3.039	84	18008	0.694	ug/L	99
22) cis-1,2-Dichloroethene	4.669	96	10256	0.405	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092623\
Data File : VU055512.D
Acq On : 26 Sep 2023 13:07
Operator : MD/SY
Sample : VIBLK142
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK142

Quant Time: Sep 27 04:48:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 27 04:44:22 2023
Response via : Initial Calibration

