

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092623\
 Data File : VU055522.D
 Acq On : 26 Sep 2023 17:26
 Operator : MD/SY
 Sample : VIBLK144
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK144

Quant Time: Sep 27 04:51:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 27 04:44:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	268592	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	275819	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	135253	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	125757	4.633	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.600%
7) Chloroethane-d5	1.911	69	96972	4.054	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.000%
11) 1,1-Dichloroethene-d2	2.563	65	52887	4.733	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	94.600%
20) 2-Butanone-d5	4.631	46	228415	46.584	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.160%
24) Chloroform-d	5.058	84	224277	4.568	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.400%
26) 1,2-Dichloroethane-d4	5.698	65	106298	4.501	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.000%
32) Benzene-d6	5.724	84	453819	4.704	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.000%
36) 1,2-Dichloropropane-d6	6.685	67	136621	4.740	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.800%
41) Toluene-d8	7.894	98	390994	4.447	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.000%
43) trans-1,3-Dichloroprop...	8.177	79	49233	4.867	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.400%
46) 2-Hexanone-d5	8.631	63	131162	42.628	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.260%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	99085	4.587	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	134112	4.913	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.200%
Target Compounds						
16) Methylene chloride	3.042	84	8201	0.365	ug/L	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092623\
Data File : VU055522.D
Acq On : 26 Sep 2023 17:26
Operator : MD/SY
Sample : VIBLK144
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK144

Quant Time: Sep 27 04:51:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 27 04:44:22 2023
Response via : Initial Calibration

